Avery Weigh-Tronix

Calibration Certificate

Avery Weigh-Tronix is a trading name of Avery Weigh-Tronix Limited
Registered Office: Foundry Lane, Smethwick, West Midlands, B66 2LP England
Registered in England No. 595129

Tel: 0845 900 22 44

Customer Details

Name : W R G
Address : Dixon House
Joseph Noble Road
Workington
CA14 4JH

Contact Name : Paul Wetherell
Telephone No. 01900602205

Certificate No. 920358478/0101

Date of Calibration : 09/12/2010
Next Calibration Due : Dec 2011

Agreement No.
Location : WORKINGTON
Department :
Plant Number :
Manufacturer : Avery Weigh-Tronix
Model : BMSE1105
Serial No : 075150431
Capacity : 50000.0 kg
Resolution : 20Kg
Tolerance : +/-60Kg
Calibration Frequency :

Test Weight Details

Weight Set No. & Certificate No.

W001 01to24: A0199

Test weights used are traceable to national standards of Mass.

Linearity / Accuracy Test

Test Point	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
0 kg			0 kg	0 kg
4,000 kg			4,000 kg	0 kg
8,000 kg			8,000 kg	0 kg
12,000 kg			12,000 kg	0 kg
16,000 kg			16,000 kg	0 kg
22,600 kg			22,600 kg	0 kg
29,920 kg			29,920 kg	0 kg
37,920 kg			37,920 kg	0 kg
44,520 kg			44,520 kg	0 kg

Off Centre Loading Test **Positions** 2 3 1 4

Test Point	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
10,600 kg			10,600 kg	0 kg
10,600 kg			10,600 kg	0 kg
10,600 kg			10,600 kg	0 kg
10,600 kg			10,600 kg	0 kg

Repeatability Test

Test Point	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
44,520 kg			44,520 kg	0 kg
44,520 kg			44,520 kg	0 kg
44,520 kg			44,520 kg	0 kg

Comments

Make-up weight used

For key calibration information please login to https://serviceportal.awtxglobal.com/ciw

Tolerances

Unless otherwise agreed tolerances will be those permitted by current Weights and Measures Regulations for Class III machines as follows:

0-500 divisions	+/- 1 divn.
>500-2000 divisions	+/- 2 divn.
>2000 divisions	+/- 3 divn.

Technicians Name: Carruthers E
Technicians Signature:

Customer Name: Paul Wetherall
Customer Signature: